



Product certificate ERNDIM IQCS Organic Acids

Product name	Control Organic Acids		
Product code	Product code ORG-02.1 ORG-02.2	Colour cap Green Red	
Date of issue	17 March 2021		
Batch numbers and expiry date	Batch number LOT 2020.1391 LOT 2020.1392	Exp. date stored at +2°C to +8°C 2025-01 2025-01	
Reconstitution volume	5.0 mL		
Estimated concentrations *	Analyte	Estimated concer	
concentrations		Level 1	Level 2
	2 Methylcitric acid	2	47
	2 OH Glutaric acid	38	394
	3 Methylglutaconic acid	16	93
	3 Methylglutaric acid	7	49
	3OH3 Methylglutaric acid	41	359
	3 OH Glutaric acid	ntd	ntd
	3 OH Isovaleric acid	9	46
	4 OH Butyric acid	12	129
	Adipic acid	18	265
	Creatinine	3100	6100
	Ethylmalonic acid	13	208
	Fumaric acid	7	217
	Glutaric acid	15	225
	Hexanoylglycine	5	36
	Isovalerylglycine	6	39
	Keto glutaric acid	31	554
	Methylmalonic acid	11	237
	Mevalonic acid	7	134
	N acetylaspartic acid	24	413
	Pyroglutamic acid	112	549
	Sebacid acid	10	84
	Suberic acid	15	173
		14	57
	Vanillacitic acid	11	86

ntd = not determined

* See ERNDIM Internal Quality Control System at the reverse



PC-ORG-02_EN.v2

Organic Acids ERNDIM IQCS

Intended purpose

These materials are control materials (thus no calibrators) for the internal control of analytical systems for the determination of organic acids in urine.

Contents

Lyophilized human urine to which organic acids have been added to achieve an analytically and physiologically relevant level of the organic acids.

Storage and stability

The product in lyophilized form is stable for 5 years when stored at $+2^{\circ}$ C to $+8^{\circ}$ C. Expiration dates are found on the product certificate (reverse). The stability of the reconstituted product is 48 hours when handled appropriately: pure water, replacement of stopper and storage at 2-8 °C.

Instructions for use

- a. Remove cap.
- b. Insert a hollow needle through the stopper to remove the vacuum (to prevent material getting on the stopper)
- c. Remove stopper.
- d. Add 5 mL aqua destillata
- e. Replace stopper
- f. Let stand for 15 minutes at room temperature
- g. Mix carefully during 20 minutes at room temperature
- h. Process product as patient sample

ERNDIM Internal Quality Control System: the Concept

The ERNDIM Internal Quality Control System (IQCS) consists of samples and a website for data management.

Samples

Samples contain analytes specifically selected for laboratories active in the field of inborn errors of metabolism. They come in two levels (1=low and 2=high) with for each analyte a relevant concentration.

Data Management

ERNDIM offers users of control materials a data management system (Note: this is an option to serve users; users do not have the obligation to use it). The strength of this system is that it does not only monitor the data of the laboratory but also compares the labs results with results of labs using the same batch of internal control materials.

In essence users can submit results every time they do an analytical run with the control material and then download two reports.

The Review Day Report shows the results of the last run in comparison to

- a) the acceptance limits set by the lab,
- b) the mean of all previous runs of the lab
- c) the mean of all laboratories.

By clicking on the name of a specific analyte in the report, Shewhart charts of that analyte are shown.

The Cumulative Table report shows the cumulative data of the lab.

Details can be found under www.erndimga.nl/General information/Use Website.

Remark

On delivery of the control materials, the certificate in the package insert shows the values as measured by a peer laboratory. Once in use laboratories submit their results and the reports will show the trimmed mean of all laboratories. This mean is a running mean which changes with every new submission: Thus a dynamic assigned value resulting from "crowd targeting".

Precautions and warnings

- 1. For in vitro diagnostic use only.
- 2. This product should be handled with care, as appropriate for biological materials. Outdated and left-over material should be discarded as potentially infectious material, according to the procedures in your institute.

References

www.ERNDIMQA.nl

Dr E.A.E. van der Hagen on behalf of the ERNDIM Internal Quality Control System Working Group

M